

# Schedule

Singapore Test Services Pte Ltd  
Block 4010 Techplace 1  
Ang Mo Kio Ave 10 #03 – 8 to 12  
Singapore 569626

Certificate No. : LA-1987-0012-C  
Issue No. : 25  
Date : 28 November 2017  
Page : 1 of 31

FIELD OF TESTING : Calibration and Measurement

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>A. MECHANICAL</b>		
<b>A.I. Pressure</b>		
Vacuum		
a. -30 in.Hg to 0 in.Hg		0.013 in.Hg
Pressure		
b. 0 inH <sub>2</sub> O to 1.0 inH <sub>2</sub> O	STS In-house Procedure T-WI-CMD-003-02 Rev 06	0.0044 in.H <sub>2</sub> O
c. 1 inH <sub>2</sub> O to 20 inH <sub>2</sub> O		0.011 in.H <sub>2</sub> O
d. 20 inH <sub>2</sub> O to 64 inH <sub>2</sub> O		0.055 in.H <sub>2</sub> O
e. 0 psi to 500 psi		0.077 psi
f. 500 psi to 2 000 psi	STS In-house Procedure T-WI-CMD-003-01 Rev 10	0.37 psi
g. 2 000 psi to 16 000 psi		8.8 psi
h. 11 psi abs to 17 psi abs	STS In-house Procedure T-WI-CMD-003-27 Rev 04	0.0025 psi abs
i. 0 psi abs to 50 psi abs		0.016 psi abs
j. 0.5 psi abs to 500 psi abs	STS In-house Procedure T-WI-CMD-003-01 Rev 10	0.091 psi abs
k. 0.5 psi abs to 2000 psi abs		0.21 psi abs
Differential Pressure		
l. - 2500 Pa to 2500 Pa	STS In-house Procedure T-WI-CMD-003-02 Rev 06	1.6 Pa

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 2 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>A.II FORCE MEASUREMENT</b></p> <p>1. Force Measurement Device (Tension/Compression)</p> <p>a. Up to 5 kgf (Tension) Up to 5 kgf (Compression)</p> <p>b. Up to 20 kgf (Tension) Up to 20 kgf (Compression)</p> <p>c. Up to 50 kgf (Tension) Up to 50 kgf (compression)</p> <p>d. Up to 500 kgf (Tension) Up to 500 kgf (Compression)</p> <p>e. Up to 10 kN (Tension) Up to 10 kN (Compression)</p> <p>f. Up to 10 000 kgf (Tension) Up to 10 000 kgf Compression)</p> <p>g. Up to 30 000 kgf (Tension) Up to 30 000 kgf(Compression)</p> <p>h. Up to 1000 kN (Compression)</p> <p>2. Platform Balance Calibration Up to 4000 kg</p> <p>3. Balances and weighing scales</p> <p>a. 0.5 to 5 g</p> <p>b. 5 to 50 g</p> <p>c. 50 to 200 g</p> <p>d. 200 to 600 g</p> <p>e. 600 to 2000 g</p> <p>f. 2 to 6 kg</p> <p>g. 6 to 10 kg</p> <p>h. 10 to 30 kg</p> <p>i. 30 to 60 kg</p> <p>j. 60 to 100 kg</p> <p>k. 100 to 150 kg</p> <p>l. 150 to 500 kg</p>	<p>STS in-house procedure T-WI-CMD-003-04 Rev 11</p> <p>STS in-house procedure T-WI-CMD-003-14 Rev 06</p> <p>STS in-house procedure T-WI-CMD-003-05 Rev 08</p>	<p>0.00084 kgf 0.00084 kgf 0.0070 kgf 0.0070 kgf 0.014 kgf 0.014 kgf 0.090 kgf 0.12 kgf 0.011 kN 0.011 kN 1.2 kgf 14.8 kgf 1.6 kN 1.6 kN 1.4 kN</p> <p>0.9 kg</p> <p>0.13 mg 0.25 mg 0.77 mg 0.0027 g 0.0076 g 0.023 g 0.035 g 0.14 g 0.0013 kg 0.013 kg 0.046 kg 0.093 kg</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 3 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>A.II FORCE MEASUREMENT</b></p> <p>4. Torque</p> <p>a. Torque Measuring Devices</p> <p>0 - 1.5 ozf.in</p> <p>0 - 10 ozf.in</p> <p>0 - 18 ozf.in</p> <p>0 - 60 cN.m</p> <p>0 - 15 kgf.cm</p> <p>0 - 5 N.m</p> <p>0 - 442.6 lbf.in</p> <p>0 - 250 lbf.ft</p> <p>b. Torque Wrench</p> <p>0.0200 N.m to 1.0000 N.m</p> <p>0.1250 N.m to 6.2500 N.m</p> <p>0.250 N.m to 12.500 N.m</p> <p>1.750 N.m to 35.000 N.m</p> <p>3.500 N.m to 70.000 N.m</p> <p>7.50 N.m to 150.00 N.m</p> <p>17.50 N.m to 350.00 N.m</p> <p>35.00 N.m to 700.00 N.m</p> <p>30.0 N.m to 1500.0 N.m</p> <p>339.0 N.m to 2500 N.m</p> <p>5. Extensometer</p> <p>a. Gauge Length up to 50 mm</p> <p>b. Classification Limit up to ASTM E83-16 Class B1 for Gauge Length 25 mm and 50 mm</p> <p>c. Strain Displacement Measurement up to 25 mm</p>	<p>BS 7882 : 2008 STS in-house procedure T-WI-CMD-003-07 Rev 07</p> <p>BS EN 6789 : 2003 STS in-house procedure T-WI-CMD-003-06 Rev 06</p> <p>STS in-house procedure T-WI-CMD-001-50 Rev 08</p> <p>STS in-house procedure T-WI-CMD-001-50 Rev 08</p> <p>ASTM E83-16</p>	<p>0.0059 ozf.in</p> <p>0.041 ozf.in</p> <p>0.076 ozf.in</p> <p>0.12 cN.m</p> <p>0.043 kgf.cm</p> <p>0.0017 N.m</p> <p>0.32 lbf.in</p> <p>0.14 lbf.ft</p> <p>0.72 % F.S.</p> <p>0.66 % F.S.</p> <p>0.67 % F.S.</p> <p>0.61 % F.S.</p> <p>0.69 % F.S.</p> <p>0.66 % F.S.</p> <p>0.59 % F.S.</p> <p>0.61 % F.S.</p> <p>0.59 % F.S.</p> <p>0.60 % F.S.</p> <p>4.2 µm</p> <p>Calibrator accuracy (12.5 mm) = 0.6 µm</p> <p>Calibrator accuracy (25 mm) = 0.6 µm</p> <p>0.000023 mm/mm</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 4 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>A.II FORCE MEASUREMENT</b></p> <p>6. Standard Weights</p> <ul style="list-style-type: none"> <li>a. 1 mg</li> <li>b. 2 mg</li> <li>c. 5 mg</li> <li>d. 10 mg</li> <li>e. 20 mg</li> <li>f. 50 mg</li> <li>g. 100 mg</li> <li>h. 200 mg</li> <li>i. 500 mg</li> <li>j. 1 g</li> <li>k. 2 g</li> <li>l. 5 g</li> <li>m. 10 g</li> <li>n. 20 g</li> <li>o. 50 g</li> <li>p. 100 g</li> <li>q. 200 g</li> <li>r. 500 g</li> <li>s. 1 kg</li> <li>t. 2 kg</li> <li>u. 5 kg</li> <li>v. 10 kg</li> <li>w. 20 kg</li> <li>x. 30 kg</li> </ul>	<p>OIML R111-1:2004 STS in-house procedure T-WI-CMD-003-08 Rev 07</p>	<ul style="list-style-type: none"> <li>0.02 mg</li> <li>0.05 mg</li> <li>0.02 mg</li> <li>0.02 mg</li> <li>0.02 mg</li> <li>0.03 mg</li> <li>0.03 mg</li> <li>0.04 mg</li> <li>0.05 mg</li> <li>0.07 mg</li> <li>0.08 mg</li> <li>0.11 mg</li> <li>0.13 mg</li> <li>0.17 mg</li> <li>0.20 mg</li> <li>0.34 mg</li> <li>0.80 mg</li> <li>0.024 g</li> <li>0.025 g</li> <li>0.029 g</li> <li>0.24 g</li> <li>0.25 g</li> <li>0.26 g</li> <li>0.24 g</li> </ul>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 5 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>A.II FORCE MEASUREMENT</b>		
7. Hardness Testing Machine		
a. Force Verification of Brinell Testing Machine Range 0 – 500 kgf Range 0 – 3000 kgf	BS EN ISO 6506-2 : 2014 EN ISO 7500-1-2015 BS EN 376 : 2011 STS in-house procedure T-WI-CMD-003-09 Rev 05	0.55 kgf 3.3 kgf
b. Force Verification of Vickers Testing Machine Range 0 – 100 kgf	BS EN ISO 6507-2 : 2005 EN ISO 7500-1-2015 BS EN 376: 2011 STS in-house procedure T-WI-CMD-003-10 Rev 05	3.3 N
c. Force Verification of Rockwell Testing Machine Range 0 – 150 kgf	BS EN ISO 6508-2 : 2015 EN ISO 7500-1-2015 BS EN 376: 2011 STS in-house procedure T-WI-CMD-003-11 Rev 05	0.17 kg
8. Tension/Compression Testing Machine		
a. 0 to 100 N (compression)		0.11 N
b. 100 to 2000 N (Tension) 100 to 2000 N (Compression)		2.3 N 2.2 N
c. 2 to 10 kN (Tension) 2 to 10 kN (Compression)	BS EN ISO 376 : 2011 BS EN ISO 7500-1 : 2015 STS In-house procedure T-WI-CMD-003-12 Rev 04	0.011 KN 0.011 KN
d. 10 to 100 KN (Tension) 10 to 100 KN ( Compression )		0.12 KN 0.12 KN
e. 100 to 1000 KN ( Tension ) 100 to 1000 KN ( Compression)		1.1 KN 1.1 KN
9. Calibration of Cable Tensiometer Up to 300 lbf	STS in-house procedure T-WI-CMD-003-17 Rev 05	1.6 lbf

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 6 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>A.II FORCE MEASUREMENT</b></p> <p>10 Calibration of Durometer (Verification of the test force)</p> <p>a. ASTM D 2240:2015 Type A b. ASTM D 2240:2015 Type D c. BS EN ISO 868:2003 Type A d. BS EN ISO 868:2003 Type D e. JIS 6301:1995 Type A f. JIS 6301:1995 Type C g. JIS 6301:1995 Type C2 h. JIS K 7215:2006 Type A</p> <p>11 Calibration of Metallic / Non Metallic Objects</p> <p>a. 0 – 210 g b. 200 g – 3 kg c. 3 kg – 30 kg</p>	<p>STS in-house procedure T-WI-CMD-003-19 Rev 10</p> <p>STS in-house procedure T-WI-CMD-003-20 Rev 06</p>	<p>0.0080 N 0.026 N 8.7 mN 51.4 mN 0.46 gf 5.1 gf 0.23 gf 0.88 gf</p> <p>0.01 g 0.02 g 0.3 g</p>
<p><b>A.III Flow Measurement</b></p> <p>1. Calibration of Fluid (Gas) Flow Measurement Instrument</p> <p>a. 0 to 300 cc/min b. 0 to 6000 cc/min c. 0 to 30000 cc/min</p> <p>2. Calibration of Fluid (Liquid) Flow Measurement Instrument</p> <p>0.3 to 50 GPM (US)</p>	<p>STS in-house procedure T-WI-CMD-003-16 Rev 05</p> <p>STS in-house procedure T-WI-CMD-003-18 Rev 05</p>	<p>4.5 cc/min 142 cc/min 400 cc/min</p> <p>0.15 % reading</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 7 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>A.IV Rotational Speed Measurement</b></p> <p>1 Calibration of Rotational Speed of Rotating Equipment</p> <p>a. 6 to 11 999 rpm</p> <p>b. 12 000 to 29 999 rpm</p> <p>c. 30 000 to 49 999 rpm</p> <p>d. 50 000 to 200 000 rpm</p> <p>2 Tachometer and RPM / Speed indicator</p> <p>a. 6 rpm to 6000 rpm</p> <p>b. 6000 rpm to 200 000 rpm</p>	<p>STS in-house procedure T-WI-CMD-004-14 Rev03</p> <p>STS in-house procedure T-WI-CMD-004-15 Rev04</p>	<p>1.5 rpm</p> <p>2.7 rpm</p> <p>5 rpm</p> <p>0.012 %</p> <p>1.6 rpm</p> <p>5.3 rpm</p>
<p><b>A.V Vibration &amp; Acoustic</b></p> <p>1. Accelerometers / Vibration</p> <p>a. 0.1 to 1000 mV/g</p> <p>b. 0.1 to 1000 pC/g</p> <p>c. 0.1 to 1000 mV/g</p> <p>d. 0.1 to 1000 pC/g</p> <p>e. 0.1 to 1000 mV/g</p> <p>f. 0.1 to 1000 pC/g</p> <p>2. Sound Level Meter</p> <p>94 dB at 1 kHz</p> <p>114 dB at 1 kHz</p>	<p>STS in-house procedure T-WI-CMD-004-16 Rev06</p> <p>10 Hz to 99 Hz</p> <p>100 Hz to 2500 Hz</p> <p>2501 Hz to 9000 Hz</p> <p>STS in-house procedure T-WI-CMD-004-20 Rev 05</p>	<p>4.2%</p> <p>4.2 %</p> <p>6.7%</p> <p>0.15dB</p> <p>0.15dB</p>



# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 8 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>B. DIMENSIONAL METROLOGY</b></p> <p><b>B.1 Limit Gauges</b></p> <p>1. Plain Plug Gauges</p> <p>a. Up to 10 mm</p> <p>b. 10 mm to 50 mm</p> <p>c. 50 mm to 150 mm</p> <p>d. 150 mm to 300 mm</p> <p>2. Plain Ring Gauges</p> <p>a. 1.5 mm to 25 mm</p> <p>b. 25 mm to 150 mm</p> <p>c. 150 mm to 300 mm</p> <p>3. Plain Gap Gauges</p> <p>a. 0.5 mm to 50 mm</p> <p>b. 50 mm to 300mm</p> <p>4. Profile Gauges</p> <p>a. Up to 200 mm</p> <p>b. Over 200 mm to 500 mm</p> <p>c. Over 500 mm to 900 mm</p> <p>d. Over 900 mm to 1400 mm</p> <p>E. Over 1400 mm to 1800 mm</p> <p>5. Calibration and inspection of Jig , Fixture, Gauges, Tools and Parts etc for Size, Height, Length, Depth, Position, Angle and Geometric Dimensioning and Tolerancing. (GD&amp;T).</p> <p>a. Up to 200 mm</p> <p>b. Over 200 mm to 500 mm</p> <p>c. Over 500 mm to 900 mm</p> <p>d. Over 900 mm to 1400 mm</p> <p>e. Over 1400 mm to 1800 mm</p> <p>6. Radius Gauge (Up to 900 mm)</p>	<p>ASME B89.1.5-1998 STS in-house procedure T-WI-CMD-001-01 Rev07</p> <p>BS 4064 : 1966, BS 4065:1966 and ASME B89.1.6-2002 STS in-house procedure T-WI-CMD-001-02 Rev08</p> <p>STS in-house procedure T-WI-CMD-001-03 Rev07</p> <p>STS in-house procedure T-WI-CMD-001-04 Rev07</p> <p>STS in-house procedure T-WI-CMD-001-05 Rev07</p> <p>STS in-house procedure T-WI-CMD-001-06 Rev07</p>	<p>0.2 µm</p> <p>0.4 µm</p> <p>0.7 µm</p> <p>0.9 µm</p> <p>0.3 µm</p> <p>0.7 µm</p> <p>1.0 µm</p> <p>2 µm</p> <p>3 µm</p> <p>1.3 µm</p> <p>2.6 µm</p> <p>4.0 µm</p> <p>5.6 µm</p> <p>6.9 µm</p> <p>1.3 µm</p> <p>2.6 µm</p> <p>4.0 µm</p> <p>5.6 µm</p> <p>6.9 µm</p> <p>4.0µm</p>



# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 9 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>B.II Measuring Instruments and Tools</b></p> <p>1. Surface Plates Size up to 2 x 2 ft Size up to 3 x 3 ft Size up to 4 x 4 ft Size up to 6.5 x 5 ft</p> <p>2. Bevel Protractors</p> <p>3. External Micrometer 0.0001mm Resolution 0.001mm Resolution</p> <p>4. Internal Micrometer</p> <p>5. Depth Micrometers Up to 300 mm</p> <p>6. Calibration of Displacement Sensor, Dial, Electrical Gauge and Indicator and Thickness Gauge Up to 30 mm Up to 100 mm Up to 600 mm</p> <p>7. Dial Test Indicators Up to 30 mm</p> <p>8. Electronic and Mechanical Height Gauges up to 300 mm up to 600 mm up to 1000 mm</p>	<p>BS 817 : 2008 STS in-house procedure T-WI-CMD-001-07 Rev 08</p> <p>BS 1685 : 2008 STS in-house procedure T-WI-CMD-001-08 Rev 08</p> <p>BS EN ISO 3611:2010 STS in-house procedure T-WI-CMD-001-09 Rev 10</p> <p>BS 959 : 2008 STS in-house procedure T-WI-CMD-001-10 Rev 06</p> <p>BS 6468 : 2008 STS in-house procedure T-WI-CMD-001-11 Rev 06</p> <p>BS EN ISO 463 : 2006 ASME B 89.1.10M-2001 STS in-house procedure T-WI-CMD-001-12 Rev 09</p> <p>BS EN ISO 9493:2010 ASME ANSI B89.1.10M-2001 STS in-house procedure T-WI-CMD-001-13 Rev 07</p> <p>BS 1643 : 2008 STS in-house procedure T-WI-CMD-001-14 Rev 07</p>	<p>1.6µm 2.1µm 2.1µm 4.4µm</p> <p>3 minute</p> <p>0.22 µm 1 µm</p> <p>1.8 µm</p> <p>1 µm</p> <p>0.4 µm 1.0 µm 2.5 µm</p> <p>0.4 µm</p> <p>0.9 µm 1.6 µm 2.6 µm</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 10 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>B.II Measuring Instruments and Tools</b>		
9. Feeler Gauges	BS 957 : 2008 STS in-house procedure T-WI-CMD-001-15 Rev 06	0.3 $\mu\text{m}$
10. Vee Blocks	BS 3731 : 1987 STS in-house procedure T-WI-CMD-001-16 Rev 06	1.8 $\mu\text{m}$
11. Engineer's Parallel	BS 906 : 1972 STS in-house procedure T-WI-CMD-001-17 Rev 05	4.1 $\mu\text{m}$
12. Engineer's Square	BS 939 : 2007 STS in-house procedure T-WI-CMD-001-18 Rev 06	2.2 $\mu\text{m}$
13. Right Angle and Box Angle Plates	BS 5535 : 1978 STS in-house procedure T-WI-CMD-001-19 Rev 06	2.2 $\mu\text{m}$
14. Straight Edge	BS 5204 Part 1 : 1975 and BS 5204 Part 2 : 1977 , STS in-house procedure T-WI-CMD-001-20 Rev 07	4.1 $\mu\text{m}$
15. Height Setting Micrometer Up to 300 mm Up to 600 mm	ISO 7863 : 1984 STS in-house procedure T-WI-CMD-001-21 Rev 07	1.0 $\mu\text{m}$ 1.8 $\mu\text{m}$
16. Steel Rule Up to 300 mm Up to 1000 mm Up to 2000 mm	JIS B 7516 : 2005 STS in-house procedure T-WI-CMD-001-22 Rev 08	1.3 $\mu\text{m}$ 6 $\mu\text{m}$ 13 $\mu\text{m}$
17. Glass Scale Up to 50mm Up to 300mm	STS in-house procedure T-WI-CMD-001-23 Rev 08	1.1 $\mu\text{m}$ 1.3 $\mu\text{m}$

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 11 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>B.II Measuring Instruments and Tools</b>		
18. Measuring Tape Up to 5 m Up to 10 m Up to 20 m Up to 30 m	JIS B 7512 : 2005 STS in-house procedure T-WI-CMD-001-24 Rev 09	16 µm 22 µm 30 µm 37 µm
19. Inside micro-Checker Up to 600mm	STS in-house procedure T-WI-CMD-001-25 Rev 05	2.8 µm
20. Caliper Checker Up to 300 mm Up to 600 mm	STS in-house procedure T-WI-CMD-001-26 Rev 05	1.3 µm 1.8µm
21. Depth Micro-Checker Up to 300 mm	STS in-house procedure T-WI-CMD-001-27 Rev 05	1.3 µm
22. Toolmaker's Straight edge	BS 852 : 1939 STS in-house procedure T-WI-CMD-001-28 Rev 05	2.0 µm
23. Bore Gauge (Three-leg Micrometer)	STS in-house procedure T-WI-CMD-001-29 Rev 07	1 µm
24. Electronic and Mechanical Calipers Up to 150 mm Up to 300 mm Up to 600 mm Up to 1000 mm Up to 2000 mm	BS EN ISO 13385-1:2011 JIS B 7507 : 2016 STS In-house procedure T-WI-CMD-001-30 Rev 08	2 µm 10 µm 10 µm 10 µm 20 µm
25. Electronic/Mechanical Depth Gauge Up to 300 mm Up to 600 mm Up to 1000 mm	BS EN ISO 13385-2-2011 STS in-house procedure T-WI-CMD-001-31 Rev 07	10 µm 10 µm 10 µm

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 12 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>B.II Measuring Instruments and Tools</b>  26. Optical Projector  27. Toolmaker Microscope  28. Universal Measuring Machine Up to 10 mm Up to 20 mm Up to 50 mm Up to 70 mm Up to 100 mm  29. Dial Gauge Calibrator 0 – 30 mm  30. Clinometer / Digital Protractor  31. Spirit Level  32. Electronic Comparator (Mu-Checker)  33. Cylinder Gauges  34. Square Master Up to 600 mm  35. Test Sieve 0.02 mm to 20 mm	JIS B 7184 : 1999 STS in-house procedure T-WI-CMD-001-32 Rev 08  JIS B 7153 : 1995  STS in-house procedure T-WI-CMD-001-33 Rev 06  STS in-house procedure T-WI-CMD-001-48 Rev 09  STS in-house procedure T-WI-CMD-001-47 Rev 07  STS in-house procedure T-WI-CMD-001-52 Rev 07  BS 958 : 1968 STS in-house procedure T-WI-CMD-001-53 Rev 06  STS in-house procedure T-WI-CMD-001-41 Rev 06  JIS B 7545 : 1982 STS in-house procedure T-WI-CMD-001-54 Rev 06  STS in-house procedure T-WI-CMD-001-55 Rev 06  ASTM E11-17 and ISO 3310-1:2016 STS in-house procedure T-WI-CMD-001-57 Rev 09	2.1 µm  2.2 µm  0.09 µm 0.11 µm 0.14 µm 0.18 µm 0.23 µm  0.3 µm  0.02 deg  0.50 arc sec  0.2 µm  0.6 µm  3.9 µm  2.3 µm

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 13 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>B.II Measuring Instruments and Tools</b>		
36. Calibration of Film Thickness And Thickness Gauge a Film Thickness up to 25mm b Thickness Gauge Up to 25mm Over 25 to 150mm	STS in-house procedure T-WI-CMD-001-64 Rev 03	0.4 $\mu\text{m}$ 3.0 $\mu\text{m}$ 15 $\mu\text{m}$
37 Linear Measurement Machine Up to 1000 mm Over 1000 to 2000 mm	STS in-house procedure T-WI-CMD-001-66 Rev 01	10 $\mu\text{m}$ 16 $\mu\text{m}$
38 Co-ordinate Measuring Machine (Contact) Measuring Linear Dimension Up to 500 mm Over 500 to 900 mm Over 900 to 1200 mm Over 1200 to 1800 mm Single and Multiple stylus contacting probing and scanning systems	ISO10360-2 : 2009 ISO10360-4 : 2000 ISO10360-5 : 2010 STS in-house procedure T-WI-CMD-001-37 Rev 07	1.0 $\mu\text{m}$ 1.7 $\mu\text{m}$ 2.2 $\mu\text{m}$ 2.6 $\mu\text{m}$
<b>III Form</b>		
1. Roundness (up to 300 mm dia.)	STS in-house procedure T-WI-CMD-001-02 Rev08	0.18 $\mu\text{m}$
<b>IV. Screw Thread Measurement</b>		
1. Parallel Screw Plug Gauge a Measurement pitches from 0.5 – 6 mm b For sizes from 1 – 150 mm	For compliance with ISO 1502, Fed Std H28 and equivalent Standards T-WI-CMD-001-59 Rev 07	2.5 $\mu\text{m}$ for sizes from 1 mm to 50 mm 3.5 $\mu\text{m}$ for sizes from 50 mm to 150mm
2. Parallel Screw Ring Gauge a Measurement pitches from 0.5 mm – 6 mm b For sized from 4 mm – 100 mm	For compliance with ISO 1502, Fed Std H 28 and equivalent standards STS In-house procedure T-WI-CMD-001-60 Rev 06	2 $\mu\text{m}$ for sizes from 4 mm to 10 mm 2.5 $\mu\text{m}$ for sizes from 10 mm to 50 mm 3.5 $\mu\text{m}$ for sizes from 50 mm to 100mm

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 14 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>C. TEMPERATURE</b></p> <p>1. Temperature Enclosure</p> <p>a -80 °C to 0 °C</p> <p>b 0 °C to 100 °C</p> <p>c 100 °C to 200 °C</p> <p>d 200 °C to 1000 °C</p> <p>2. PT 100 Sensor</p> <p>a -80 °C to 0 °C</p> <p>b 0 °C to 50 °C</p> <p>c 50 °C to 200 °C</p> <p>3. Temperature Indicator c/w Sensor PT100 / Thermistor</p> <p>a -80 °C to 0 °C</p> <p>b 0 °C to 50 °C</p> <p>c 50 °C to 200 °C</p> <p>Base Metal: Type E, K, J, N &amp; T Noble Metal: Type B, R &amp; S</p> <p>a -80 °C to 0 °C</p> <p>b 0 °C to 100 °C</p> <p>c 100 °C to 260 °C</p> <p>d 260 °C to 500 °C</p> <p>e 500°C to 1000 °C</p> <p>4. Liquid in Glass Thermometer Graduation 0.1°C</p> <p>a 0 °C to 200 °C (Partial Immersion)</p> <p>b 0 °C to 200 °C (Total Immersion)</p> <p>5. Calibration Bath - Liquid Bath</p> <p>a -80 °C to 260 °C</p> <p>6. Temperature Block Calibrator</p> <p>a. -45 °C to 140 °C</p> <p>b. 140 °C to 650 °C</p> <p>c. 650°C to 1000 °C</p>	<p>STS in-house procedure T-WI-CMD-002-01 Rev.16</p> <p>STS in-house procedure T-WI-CMD-002-03 Rev 13</p> <p>STS in-house procedure T-WI-CMD-002-05 Rev 15</p> <p>STS in-house procedure T-WI-CMD-002-07 Rev 13</p> <p>STS in-house procedure T-WI-CMD-002-02 Rev 12</p> <p>STS In-house procedure T-WI-CMD-002-13 Rev 07</p>	<p>1.5 °C</p> <p>1.4°C</p> <p>1.6°C</p> <p>2.1 °C</p> <p>0.43°C</p> <p>0.11 °C</p> <p>0.37 °C</p> <p>0.43 °C</p> <p>0.11 °C</p> <p>0.37 °C</p> <p>0.43 °C</p> <p>0.11 °C</p> <p>0.39°C</p> <p>2.6 °C</p> <p>3.7 °C</p> <p>0.6 °C</p> <p>0.59 °C</p> <p>0.68°C</p> <p>0.86 °C</p> <p>3.7 °C</p> <p>4.1°C</p>





# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 16 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
8b Temperature Simulator (Source)  i. Type K -200 °C to 1300 °C  ii. Type T -200 °C to 400 °C  iii. Type J -200 °C to 1200 °C  iv. Type N -200 °C to 1300 °C  v. Type E -200 °C to 1000 °C  vi. Type R 0 °C to 1700 °C  vii. Type S 0 °C to 1700 °C  viii RTD 100Ω (Pt(385)) -200 °C to 800 °C	STS in-house procedure T-WI-CMD-002-06 Rev 10	0.10 °C  0.09 °C  0.18 °C  0.14 °C  0.13 °C  0.72 °C  0.72 °C  0.079 °C
<b>C. TEMPERATURE</b>		
9. Thermo-Hygrometer/Thermo- Hygrogaph a. 20 °C to 50 °C  b. 30 % rh to 95 % rh	STS in-house procedure T-WI-CMD-002-08 Rev 10	0.18°C  2.0% relative humidity

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 17 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>10. Humidity Chamber</p> <p>a 30 % rh at 20 °C</p> <p>b 30 % rh to 50 % rh at 20 °C</p> <p>c 50 % rh to 70 % rh at 20 °C</p> <p>d 70 % rh to 95 % rh at 20 °C</p> <p>e 30 % rh at 30 °C</p> <p>f 30 % rh to 50 % rh at 30 °C</p> <p>g 50 % rh to 70 % rh at 30 °C</p> <p>h 70 % rh to 95 % rh at 30 °C</p> <p>i 30 % rh at 50 °C</p> <p>j 30 % rh to 50 % rh at 50 °C</p> <p>k 50 % rh to 70 % rh at 50 °C</p> <p>l 70 % rh to 95 % rh at 50 °C</p> <p>% rh - % relative humidity</p>	<p>STS in-house procedure T-WI-CMD-002-09 Rev 11</p>	<p>1.4 % relative humidity</p> <p>2.1 % relative humidity</p> <p>2.9 % relative humidity</p> <p>4.1 % relative humidity</p> <p>1.2 % relative humidity</p> <p>1.7 % relative humidity</p> <p>2.3 % relative humidity</p> <p>3.8 % relative humidity</p> <p>1.9% relative humidity</p> <p>1.9% relative humidity</p> <p>2.6 % relative humidity</p> <p>2.9 % relative humidity</p>
<p><b>D. ELECTRICAL</b></p> <p>1. DC Voltage Measuring Instrument</p> <p>a. 0.01mV to 0.09999mV</p> <p>b. 0.1mV to 0.99999mV</p> <p>c. 1mV to 9.99999mV</p> <p>d. 10mV to 99.99999mV</p> <p>e. 100mV to 199.99999mV</p> <p>f. 200mV to 219.99999mV</p> <p>g. 0.22V to 2.1999999V</p> <p>h. 2.2V to 10.999999V</p> <p>i. 11V to 21.999999V</p> <p>j. 22V to 219.99999V</p> <p>k. 220V to 1100V</p>	<p>STS In-house procedure T-WI-CMD-004-03 Rev 08</p>	<p>4.2%</p> <p>0.42%</p> <p>0.042%</p> <p>50 ppm</p> <p>12 ppm</p> <p>9.6 ppm</p> <p>8.3 ppm</p> <p>4.7 ppm</p> <p>4 ppm</p> <p>6.9 ppm</p> <p>8.4 ppm</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 18 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>2 AC Voltage Measuring Instrument</p> <p>a. 0.50001 mV - 1 mV</p> <p>b. 1.000001 mV - 2.19999 mV</p> <p>c. 2.2 mV - 5 mV</p>	<p>STS In-house procedure T-WI-CMD-004-03 Rev 08</p> <p>10Hz to 20Hz 20Hz to 40Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz 300kHz to 500kHz</p> <p>10Hz to 20Hz 20Hz to 40Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz 300kHz to 500kHz 500kHz to 1MHz</p> <p>10Hz to 20Hz 20Hz to 40Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz 300kHz to 500kHz 500kHz to 1MHz</p>	<p>0.9%</p> <p>0.9%</p> <p>0.9%</p> <p>1.1%</p> <p>1.1%</p> <p>2.4 %</p> <p>4.2 %</p> <p>0.43 %</p> <p>0.43 %</p> <p>0.42 %</p> <p>0.43 %</p> <p>0.56 %</p> <p>1.2 %</p> <p>2.2 %</p> <p>2.3 %</p> <p>0.21 %</p> <p>0.2 %</p> <p>0.21 %</p> <p>0.21 %</p> <p>0.28 %</p> <p>0.56 %</p> <p>1.1 %</p> <p>1.2 %</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 19 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b>		
2 AC Voltage Measuring Instrument		
d. 5.00001 mV - 10 mV	10Hz to 20Hz 20Hz to 40Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz 300kHz to 500kHz 500kHz to 1MHz	0.11 % 0.09 % 0.092 % 0.11 % 0.16 % 0.31 % 0.54 % 0.68 %
e. 10.00001 mV - 21.99999 mV	10Hz to 20Hz 20Hz to 40Hz 40Hz to 20kHz 20kHz to 50kHz 50kHz to 100kHz 100kHz to 300kHz 300kHz to 500kHz 500kHz to 1MHz	0.065 % 0.052 % 0.05 % 0.061 % 0.11 % 0.21 % 0.35 % 0.48 %
f. 22 mV - 50 mV	10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 300 kHz 300 kHz - 500 kHz 500 kHz - 1 MHz	0.059 % 0.049 % 0.029 % 0.041 % 0.095 % 0.16 % 0.24 % 0.4 %
g. 50.00001 mV - 100 mV	10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 300 kHz 300 kHz - 500 kHz 500 kHz - 1 MHz	0.054 % 0.049 % 0.024 % 0.035 % 0.14% 0.14 % 0.2 % 0.37 %

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 20 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b> 2 AC Voltage Measuring Instrument h. 0.1V - 0.21999999 V  i. 0.22V - 0.5 V  j. 0.5000001 V - 1 V  k. 1.00001 V - 2.1999999 V	10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 300 kHz 300 kHz - 500 kHz 500 kHz - 1 MHz  10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 300 kHz 300 kHz - 500 kHz 500 kHz - 1 MHz  10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 300 kHz 300 kHz - 500 kHz 500 kHz - 1 MHz  10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 300 kHz 300 kHz - 500 kHz 500 kHz - 1 MHz	0.037 % 0.037 % 0.017 % 0.028 % 0.064 % 0.12 % 0.32 % 0.32 %  0.045 % 0.043 % 0.017 % 0.013 % 0.025 % 0.079 % 0.2 % 0.31 %  0.033 % 0.033 % 0.013 % 96 ppm 0.018 % 0.058 % 0.14 % 0.24 %  0.029 % 0.029 % 0.011 % 85 ppm 0.015 % 0.051 % 0.13 % 0.21 %

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 21 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>D. ELECTRICAL</b></p> <p>2 AC Voltage Measuring Instrument</p> <p>l. 2.2 V - 5 V</p> <p>m. 5.000001 V - 10 V</p> <p>n. 10.000001 V - 21.999999 V</p> <p>o. 22 V - 50 V</p>	<p>10 Hz - 20 Hz</p> <p>20 Hz - 40 Hz</p> <p>40 Hz - 20 kHz</p> <p>20 kHz - 50 kHz</p> <p>50 kHz - 100 kHz</p> <p>100 kHz - 300 kHz</p> <p>300 kHz - 500 kHz</p> <p>500 kHz - 1 MHz</p> <p>10 Hz - 20 Hz</p> <p>20 Hz - 40 Hz</p> <p>40 Hz - 20 kHz</p> <p>20 kHz - 50 kHz</p> <p>50 kHz - 100 kHz</p> <p>100 kHz - 300 kHz</p> <p>300 kHz - 500 kHz</p> <p>500 kHz - 1 MHz</p> <p>10 Hz - 20 Hz</p> <p>20 Hz - 40 Hz</p> <p>40 Hz - 20 kHz</p> <p>20 kHz - 50 kHz</p> <p>50 kHz - 100 kHz</p> <p>100 kHz - 300 kHz</p> <p>300 kHz - 500 kHz</p> <p>500 kHz - 1 MHz</p> <p>10 Hz - 20 Hz</p> <p>20 Hz - 40 Hz</p> <p>40 Hz - 20 kHz</p> <p>20 kHz - 50 kHz</p> <p>50 kHz - 100 kHz</p> <p>100 kHz - 300 kHz</p>	<p>0.043 %</p> <p>0.043 %</p> <p>0.016 %</p> <p>0.013 %</p> <p>0.02 %</p> <p>0.055 %</p> <p>0.2 %</p> <p>0.3 %</p> <p>0.033 %</p> <p>0.033 %</p> <p>0.013 %</p> <p>96 ppm</p> <p>0.015 %</p> <p>0.04 %</p> <p>0.15 %</p> <p>0.22 %</p> <p>0.029 %</p> <p>0.029 %</p> <p>0.011 %</p> <p>86 ppm</p> <p>0.013 %</p> <p>0.034 %</p> <p>0.13 %</p> <p>0.19 %</p> <p>0.043 %</p> <p>0.043 %</p> <p>0.016 %</p> <p>0.013 %</p> <p>0.027 %</p> <p>0.17 %</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 22 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b> 2 AC Voltage Measuring Instrument  p. 50.00001 V - 100 V  q. 100.00001 V - 219.99999 V  r. 220 V - 249.9999 V  s. 220 V - 500 V  t. 501 V - 1000 V	10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz  10 Hz - 20 Hz 20 Hz - 40 Hz 40 Hz - 20 kHz 20 kHz - 50 kHz 50 kHz - 100 kHz  15 Hz - 50 Hz  50 Hz - 1 kHz  50 Hz - 1 kHz	0.033 % 0.033 % 0.013 % 0.011 % 0.021 %  0.031 % 0.031 % 0.019 % 91 ppm 0.018 %  0.041 %  0.032 %  78 ppm
3 DC Current Measuring Instrument a. 0.1uA to 0.9999uA b. 1uA to 9.9999uA c. 10uA to 99.9999uA d. 100uA to 199.9999uA e. 200uA to 219.9999uA f. 0.22mA to 2.19999mA g. 2.2mA to 21.9999mA h. 22mA to 219.9999mA i. 0.22A to 2.2A j. 2.2A to 10 A k. 10.1A to 550A l. 551A to 1000A	STS In-house procedure T-WI-CMD-004-03 Rev 08	6.1 % 0.61 % 0.065 % 0.011 % 71 ppm 67 ppm 40 ppm 53 ppm 0.039 % 0.059 % 0.79 % 0.82 %



# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 23 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b> 4 AC Current Measuring Instrument a. 9 $\mu$ A – 99.999 $\mu$ A  b. 100 $\mu$ A – 199.999 $\mu$ A  c. 200 $\mu$ A – 219.999 $\mu$ A  d. 0.22 mA – 2.19999 mA  e. 2.2 mA – 21.9999 mA  f. 22 mA – 219.999 mA	T-WI-CMD-004-03 Rev 08  10Hz to 20Hz 20Hz to 40Hz 40Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz  10Hz to 20Hz 20Hz to 40Hz 40Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz  10Hz to 20Hz 20Hz to 40Hz 40Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz  10Hz to 20Hz 20Hz to 40Hz 40Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz  10Hz to 20Hz 20Hz to 40Hz 40Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz	0.21 % 0.21 % 0.13 % 0.17 % 0.84 %  0.045 % 0.045 % 0.03 % 0.041 % 0.18 %  0.037 % 0.037 % 0.023 % 0.035 % 0.15 %  0.036 % 0.036 % 0.028 % 0.034 % 0.18 %  0.034 % 0.034 % 0.027 % 0.026 % 0.17 %  0.034 % 0.034 % 0.027 % 0.024 % 0.14 %

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 24 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>D. ELECTRICAL</b></p> <p>4 AC Current Measuring Instrument</p> <p>g. 0.22 A – 2.19999 A</p> <p>h. 2.2 A – 20 A</p> <p>i. 20 A – 550 A</p> <p>j. 551 A – 1000 A</p> <p>5 Resistance Measuring Instrument Range Values</p> <p>a. 1Ω to 10.99Ω</p> <p>b. 11Ω to 32.999Ω</p> <p>c. 33Ω to 109.999Ω</p> <p>d. 110Ω to 329.999Ω</p> <p>e. 0.33kΩ to 1.09999kΩ</p> <p>f. 1.1kΩ to 3.29999kΩ</p> <p>g. 3.3kΩ to 10.9999kΩ</p> <p>h. 11kΩ to 32.9999kΩ</p> <p>i. 33kΩ to 109.999kΩ</p> <p>j. 110kΩ to 329.999kΩ</p> <p>k. 0.33MΩ to 0.99999MΩ</p> <p>l. 1MΩ to 10MΩ</p> <p>m. 10MΩ to 15.9999MΩ</p> <p>n. 16MΩ to 20MΩ</p> <p>o. 20MΩ to 32.9999MΩ</p> <p>p. 33MΩ to 1.1GΩ</p> <p>q. 1.2GΩ to 111GΩ</p>	<p>20Hz to 5kHz</p> <p>45Hz to 1kHz</p> <p>45Hz to 65Hz</p> <p>40Hz to 60Hz</p> <p>STS In-house procedure T-WI-CMD-004-03 Rev 08</p>	<p>0.69 %</p> <p>0.33 %</p> <p>0.71 %</p> <p>0.6 %</p> <p>0.64 %</p> <p>0.12 %</p> <p>0.043%</p> <p>0.018 %</p> <p>0.021%</p> <p>0.012 %</p> <p>0.022 %</p> <p>0.012 %</p> <p>0.023 %</p> <p>0.014 %</p> <p>0.025 %</p> <p>0.057 %</p> <p>0.081 %</p> <p>0.081 %</p> <p>0.081 %</p> <p>1.3 %</p> <p>2.4 %</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 25 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>D. ELECTRICAL</b></p> <p>6 Resistance Measuring Instrument Fixed values</p> <p>a. 1 <math>\Omega</math>, 1.9 <math>\Omega</math></p> <p>b. 10 <math>\Omega</math>, 19 <math>\Omega</math></p> <p>c. 100 <math>\Omega</math>, 190 <math>\Omega</math></p> <p>d. 1 k<math>\Omega</math>, 1.9 k<math>\Omega</math>, 10 k<math>\Omega</math>, 19 k<math>\Omega</math></p> <p>e. 100 k<math>\Omega</math>, 190 k<math>\Omega</math></p> <p>f. 1 M<math>\Omega</math></p> <p>g. 1.9 M<math>\Omega</math></p> <p>h. 10 M<math>\Omega</math></p> <p>i. 19 M<math>\Omega</math></p> <p>j. 100 M<math>\Omega</math></p> <p>7 Capacitance Measuring Instrument Range Values</p> <p>a. 0.33nF to 10.999nF</p> <p>b. 11nF to 109.99nF</p> <p>c. 110nF to 329.99nF</p> <p>d. 0.33uF to 1.0999uF</p> <p>e. 1.1uF to 3.2999uF</p> <p>f. 3.3uF to 10.999uF</p> <p>g. 11uF to 32.999uF</p> <p>h. 33uF to 109.99uF</p> <p>i. 110uF to 329.99uF</p> <p>j. 0.33mF to 1.1mF</p> <p>8 Capacitance Measuring Instrument Fixed Values</p> <p>a. 1 pF</p> <p>b. 10 pF</p> <p>c. 100 pF</p> <p>d. 1000 F</p> <p>9 Inductance Measuring Instrument Ranged values</p> <p>a 1 mH - 1000 mH</p>	<p>STS In-house procedure T-WI-CMD-004-03 Rev 08</p> <p>STS In-house procedure T-WI-CMD-004-03 Rev 08</p> <p>STS In-house procedure T-WI-CMD-004-03 Rev 08</p> <p>STS In-house procedure T-WI-CMD-004-13 Rev 03</p>	<p>97 ppm</p> <p>29 ppm</p> <p>18 ppm</p> <p>14 ppm</p> <p>15 ppm</p> <p>21 ppm</p> <p>25 ppm</p> <p>44 ppm</p> <p>56 ppm</p> <p>0.013 %</p> <p>58 pF</p> <p>0.33 nF</p> <p>0.43 nF</p> <p>0.0033 uF</p> <p>0.013 uF</p> <p>0.044 uF</p> <p>0.15 uF</p> <p>0.58 uF</p> <p>2.4 uF</p> <p>0.010 mF</p> <p>0.13 %</p> <p>0.12 %</p> <p>0.12 %</p> <p>0.12 %</p> <p>1.2 %</p>

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 26 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b>		
10 DC Voltage Source	STS In-house procedure	
a. 0.01mV to 0.1mV	T-WI-CMD-004-09 Rev 08	3.6%
b. 0.10001mV to 1mV	T-WI-CMD-004-06 Rev 04	0.37%
c. 1.00001mV to 10mV		0.037%
d. 10.00001mV to 100mV		49 ppm
e. 100.00001mV to 120mV		16 ppm
f. 0.12000001V to 1.2V		15 ppm
g. 1.2000001V to 12V		17 ppm
h. 12.000001V to 120V		17 ppm
i. 120.00001V to 1000V		20 ppm
j. 1000V to 2000V		0.091%
k. 2000V to 15000V		0.28%
11 AC Voltage Source	STS In-house procedure	
	T-WI-CMD-004-09 Rev 08	
	T-WI-CMD-004-06 Rev 04	
a. 1.0001mV - 12 mV	40 Hz - 1 kHz	0.39 %
	1 kHz - 20 kHz	0.17 %
	20 kHz - 50 kHz	0.25 %
	50 kHz - 100 kHz	0.71 %
b. 12.0001mV - 120 mV	40 Hz - 1 kHz	0.047 %
	1 kHz - 20 kHz	0.036 %
	20 kHz - 50 kHz	0.055 %
	50 kHz - 100 kHz	0.12 %
	100 kHz - 300 kHz	0.45 %
c. 0.120001V - 1.2 V	40 Hz - 1 kHz	0.047 %
	1 kHz - 20 kHz	0.036 %
	20 kHz - 50 kHz	0.055 %
	50 kHz - 100 kHz	0.12 %
	100 kHz - 300 kHz	0.45 %
d. 1.20001 V - 12 V	40 Hz - 1 kHz	0.062 %
	1 kHz - 20 kHz	0.043 %
	20 kHz - 50 kHz	0.06 %
	50 kHz - 100 kHz	0.16 %
e. 12.0001 V - 120 V	40 Hz - 1 kHz	0.063 %
	1 kHz - 20 kHz	0.043 %

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 27 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b>		
f. 120.001 V - 700 V	50 Hz - 1 kHz 1 kHz - 20 kHz	0.085 % 0.089 %
11 AC Voltage Source		
g. 700 V to 750 V	50 Hz to 1 kHz	0.13 %
h. 750 V to 2000 V	50 Hz to 1 kHz	0.13 %
i. 2000 V to 10000 V	50Hz	2.3 %
12 DC Current Source	STS In-house procedure	
a. 1nA to 10nA	T-WI-CMD-004-09 Rev 08	4.7 %
b. 10.001nA to 100nA	T-WI-CMD-004-06 Rev 04	0.46 %
c. 110nA to 120nA		0.047 %
d. 0.1201uA to 1.2uA		0.043 %
e. 1.2001uA to 12uA		0.013 %
f. 12.0001uA to 120uA		0.011 %
g. 0.1200001mA to 1.2mA		78 ppm
h. 1.200001mA to 12mA		95 ppm
i. 12.00001mA to 120mA		95 ppm
j. 0.1200001A to 1.05A		0.023 %
k. 1.20001A to 3A		0.18 %
l. 3A to 3.6A		0.21 %
m. 3.60001A to 10 A		0.2 %
13 AC Current Source	STS In-house procedure	
	T-WI-CMD-004-09 Rev 08	
	T-WI-CMD-004-06 Rev 04	
a. 10.0001μA - 120 μA	10Hz to 20Hz 20Hz to 45Hz 45Hz to 100Hz 100Hz to 1kHz 1kHz to 5 kHz 5kHz to 10 kHz	0.81 % 0.81 % 0.53 % 0.42 % 0.42 % 2.4 %
b. 0.12001mA – 1.2 mA	10Hz to 20Hz 20Hz to 45Hz 45Hz to 100Hz 100Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz	0.66 % 0.66 % 0.37 % 0.27 % 0.23 % 0.27 %

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 28 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b>		
c. 1.20001 mA – 12 mA	10Hz to 20Hz 20Hz to 45Hz 45Hz to 100Hz 100Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz	0.66 % 0.66 % 0.37 % 0.27 % 0.23 % 0.27 %
d. 12.0001 mA – 120 mA	10Hz to 20Hz 20Hz to 45Hz 45Hz to 100Hz 100Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz	0.66 % 0.66 % 0.38 % 0.31 % 0.31 % 0.54 %
e. 0.120001 A – 1.00000 A	10Hz to 20Hz 20Hz to 45Hz 45Hz to 100Hz 100Hz to 1kHz 1kHz to 5kHz 5kHz to 10kHz	0.66 % 0.66 % 0.38 % 0.31 % 0.31 % 0.54 %
f. 1.00001 A – 3.00000 A	45Hz to 1kHz	0.67%
g. 3.00001 A – 3.60000 A	45Hz to 1 kHz	0.7 %
h. 3.60001 A – 10 A	45Hz to 1kHz	0.64%
14 Resistance Source	STS In-house procedure	
a. 0.001Ω to 0.01Ω	T-WI-CMD-004-09 Rev 08	6.4 %
b. 0.011Ω to 0.1Ω	T-WI-CMD-004-06 Rev 04	0.62 %
c. 0.11Ω to 10Ω		0.068 %
d. 10.001Ω to 100Ω		76 ppm
e. 100.001Ω to 1kΩ		0.06 %
f. 1.001kΩ to 10kΩ		17 ppm
g. 10.001kΩ to 100kΩ		21 ppm
h. 100.001kΩ to 1MΩ		44 ppm
i. 1.00001MΩ to 12MΩ		0.068 %
j. 12.0001MΩ to 120MΩ		0.068 %
k. 0.121GΩ to 1GΩ		0.067 %

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 29 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b>		
15 Capacitance Source	STS In-house procedure T-WI-CMD-004-09 Rev 08	
a. 1nF to 10nF	1 Vac @ 1kHz	0.12 %
b. 10.0001nF to 100nF	1 Vac @ 1kHz	0.12 %
c. 100.001nF to 1000nF	1 Vac @ 1kHz	0.12 %
d. 1.0001uF to 10uF	1 Vac @ 1kHz	0.35 %
16 Frequency Measuring Instrument 100mHz to 1.3GHz	STS In-house procedure T-WI-CMD-004-03 Rev 08	2.4 ppm
17 Stopwatch / Timer	STS In-house procedure T-WI-CMD-004-10 Rev 06	0.13s
a. 1s to 24hr		
18 Frequency and Time standard (RF)	STS In-house procedure	
a. 5 MHz	T-WI-CMD-004-04 Rev 06	$3.9 \times 10^{-10}$
b. 10 MHz	T-WI-CMD-004-09 Rev 08	$4.1 \times 10^{-10}$
c. 100 kHz – 20 GHz		$2.4 \times 10^{-9}$
19 RF Power	STS In-house procedure T-WI-CMD-004-04 Rev 05	
a. 20 dBm to -20 dBm at 1 GHz		0.17 dB
b. 10 dBm to -20 dBm at 10 GHz		0.36 dB
c. 10 dBm to -20 dBm at 18 GHz		0.36 dB
20 Amplitude Modulation	STS In-house procedure T-WI-CMD-004-04 Rev 05	
5% to 95% depth	Carrier Frequency at	
50 Hz – 10 kHz	150 kHz – 10 MHz	2.7 %
50 Hz – 50 kHz	10 MHz – 1.3 GHz	2.1 %
50 Hz – 50 kHz	1.3 GHz – 10 GHz	2.1 %
21. Frequency Modulation	STS In-house procedure T-WI-CMD-004-04 Rev 05	
	Carrier Frequency at	
a. 20 Hz - 10 kHz	250 kHz – 10 MHz, dev 40kHz	3 %
b. 50 Hz - 100 kHz	10 MHz – 1.3 GHz, dev 400kHz	1.2 %
c. 20 Hz - 200 kHz	10 MHz – 1.3 GHz, dev 400kHz	5.9 %
d. 50 Hz - 100 kHz	10 MHz – 18 GHz, dev 100kHz	1.3 %



# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 30 of 31

MEASURED QUANTITIES/ RANGE/ INSTRUMENTS TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>D. ELECTRICAL</b>		
22 Phase Modulation (Modulation rate with peak phase deviation)	STS In-house procedure T-WI-CMD-004-04 Rev 05 Carrier Frequency at 150 kHz- 10 MHz	5 %
a. 200 Hz - 10 kHz @ up to 4 radians		
b. 200 Hz - 20 kHz @ up to 20 radians	10 MHz – 1.3 GHz	4.1 %
c. 200 Hz - 20 kHz @ up to 20 radians	1.3 GHz – 20 GHz	4.1 %
23 Oscilloscope	STS In-house procedure T-WI-CMD-004-11 Rev 05	
a. Vertical Deflection  2mV – 30V		2 %
b. Horizontal Deflection 2ns – 4s		0.15 %
c. Bandwidth (-3dB) 50 kHz – 500 MHz		4.6 %
<b>E OPTICAL</b>		
1. Light Meter	STS in-house procedure T-WI-CMD-004-18 Rev 04	
10 lx – 99 lx		6.4%
100 lx – 4000 lx		6.1%

\* CMC is expressed as an expanded uncertainty estimated at a level of confidence of approximately 95%.

# Schedule



Certificate No. : LA-1987-0012-C

Issue No. : 25

Date : 28 November 2017

Page : 31 of 31

## Approved signatories

Name	Cat. A Mechanical	Cat. B Dimensional	Cat. C Temperature	Cat. D Electrical	Cat. E Optical
Mr Mohd Haron B. Junaidi	A(I), A(II 1 to 3)	All	-	-	-
Mr Law Eng Soon	A(I), A(II 1 to 5 )	All	-	-	-
Mr Pek Leong Beng, Joseph	A(I), A(II), A(III)	-	All	-	-
Mr Toh Kheng Leong	A(I), A(II – excluding 5,7 to 8), A(III 2)	-	-	-	-
Mr Yap Hwee Soon	A(I), A(II), A(III), A(IV),A(V1).	-	All	All- excluding 7,8, 18 to 22	-
Mr Wong Sai Kee, Frankie	A(IV)	-	C ( excluding 4 & 7)	1, 2, 3a to 3j, 4a to 4h, 5 to 9, 16, 17	-
Mr Ang Zhong Li	-	-	-	1, 2, 3a to 3j, 4a to 4h, 5	-
Mr Yip Hang Ming, Sam	A(V)	-	-	All	All
Mr Lee Boon Chong, Alvin	-	(I)1, 2, 5, (II)1, 3 to 11, 15 to 18, 20, 21, 23 to 28, 32, 33, 36, 37, 38, (III)1.	-	-	-
Ms Ong Yok Khuan	-	(I) 1,2,5 (II)3,6 to 8, 24	-	-	-
Mr Teo Tong Nee, Anthony	-	(I) 1,2,5 (II)3,6 to 8, 24	-	-	-

## Note :

This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. A laboratory's fulfilment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid test results. The **management system requirements** in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001.